



- NOTES:
- DETECTOR SHOULD BE MOUNTED AS FOLLOWS:
 - THE DETECTOR SCINTILLATOR SHOULD BE SET AT THE DETECTOR GAP RECOMMENDED ON THE SIZE SHEET. IF A SIZE SHEET IS NOT AVAILABLE, VEGA RECOMMENDS 50mm [2"] FOR VESSELS WITH MAXIMUM SURFACE TEMPERATURE LESS THAN 50°C AND AT LEAST 150mm [6"] FOR VESSELS WITH MAXIMUM SURFACE TEMPERATURE GREATER THAN 50°C. THIS SPACING IS FROM THE VESSEL'S EXTREME OUTER SURFACE (INSULATION, SHEATHING, ETC.) TO THE SCINTILLATOR SURFACE.
 - DO NOT REMOVE ANY INSULATION FROM THE VESSEL DURING INSTALLATION.
 - DO NOT PLACE ANY MATERIAL BETWEEN THE VESSEL'S EXTREME OUTER SURFACE AND THE SCINTILLATOR UNLESS DIRECTED BY VEGA.
 - MAKE SURE PROCESS MATERIAL OR OTHER FOREIGN MATTER CANNOT ACCUMULATE BETWEEN VESSEL AND SCINTILLATOR.
 - EXPOSED MATERIAL IS 316 STAINLESS STEEL.
 - APPROXIMATE WEIGHT OF BRACKET IS 2.8 kg [6.2 lb].
APPROXIMATE WEIGHT OF CONDUIT HANGER IS .14 kg [.3 lb] EACH.

Projektionsmethode Projection angle		Achtung: Diese Zeichnung stellt die zum Zeitpunkt der Konfiguration gültige Geräteausführung dar. Sie wurde automatisch erstellt, spätere Konfiguration können hiervon abweichen. Attention: This drawing shows the valid instrument version on the configuration date. It was created automatically, later configurations may vary.
Maßstab Scale	1:7	eID MB4997-US-PC-4
Originalformat size of origin	B	Produktschlüssel Model Code FIBERTRAC 31 PTB3 MOUNTING
	 www.vega.com	Datum/Date 2019-09-26
		Alle Bemaßungen in mm [inch] / All dimensions are mm [inches]